

## Unjointed Finger Probe -- UFP25



This Unjointed Finger Probe is made for forceful insertion where the flexibility of the Jointed Finger Probe would be difficult. The UFP25 meets IEC 61032 requirements, Figure 7, Test Probe 11. The “finger” portion of the probe is made of stainless steel and the rest of the instrument of Delrin®. The handle is designed with a threaded hole for use with force gauges.