





Jewelry Probe, 3mm Diameter -- JP03



The **3 mm Jewelry Probe** tests to IEC 61010-1, Clause 6.2.3 and others. Material is stainless steel - scribed at 100 mm

Jewelry Probe, 4mm Diameter – Model JP04



The **4 mm Jewelry Probe** tests to IEC 61010-1, Clause 6.2.2.; IEC 60601, Clause 16(6). String handle is used to hold probe in vertical position for testing. Material is stainless steel - scribed at 100 mm.



PO Box 964, Southampton, PA 18966

© 2001- 2003 Ergonomics Inc. - all rights reserved

Phone: 800-862-0102, 215-357-5124

Fax: 215-364-7582

E-Mail: info@ergonomicsusa.com

Website: www.ergonomicsusa.com